



"WE MAKE NGS BETTER"

cfKapture Kit - DX (1-2 mL)

Isolation of cfDNA from Plasma

Catalog Nos. CFK-DI5-2MLE, CFK-DI50-2MLE
Manual Revision 0
WI-72-149


- Isolation of circulating cell-free DNA from plasma
- Magnetic bead-based chemistry

Instructions For Use

Contents

Product Description and Process	1
Kit Contents, Shipping and Storage, and Safety Info	1
Preparation of Reagents	2
Protocol: 1 mL sample volume (15 mL tube format)	3
Protocol: 2 mL sample volume (15 mL tube format)	5
Troubleshooting Guide	7
Ordering information	8

EC REP CEpartner4U
 Esdoornlaan 13, 3951 DB Maarn
 Netherlands
 www.cepartner4u.com

 MagBio Genomics, Inc.
 200 Professional Drive
 Gaithersburg, MD 20879
 USA



For *in vitro* diagnostic procedures.

Information in this document is subject to change without notice.

MAGBIO GENOMICS, INC. DISCLAIMS ALL WARRANTIES WITH RESPECT TO THIS DOCUMENT, EXPRESSED OR IMPLIED, INCLUDING BUT NOT LIMITED TO THOSE OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. TO THE FULLEST EXTENT ALLOWED BY LAW, IN NO EVENT SHALL MAGBIO GENOMICS, INC. BE LIABLE, WHETHER IN CONTRACT, TORT, WARRANTY, OR UNDER ANY STATUTE OR ON ANY OTHER BASIS FOR SPECIAL, INCIDENTAL, INDIRECT, PUNITIVE, MULTIPLE OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH OR ARISING FROM THIS DOCUMENT, INCLUDING BUT NOT LIMITED TO THE USE THEREOF, WHETHER OR NOT FORESEEABLE AND WHETHER OR NOT MAGBIO GENOMICS, INC. IS ADVISED OF THE POSSIBILITY OF SUCH DAMAGES.

TRADEMARKS


The trademarks mentioned herein are the property of MagBio Genomics, Inc. or their respective owners.

Product Description

The cfKapture Kit - DX is used for purification of circulating, cell-free nucleic acids from plasma. It's designed for the purification of circulating cell-free DNA (ccfDNA) from maternal and cancer patient's blood plasma. The kit can isolate and concentrate cell-free DNA from fresh or frozen samples. It eliminates high molecular weight DNA, leaving behind cfDNA. The isolated cfDNA can be directly used for real time-PCR and DNA library preparation suitable for next generation sequencing. This protocol can be used in manual procedures as well as a guideline for adapting the kit to automated instruments.

This product is intended to be used by qualified and trained laboratory professionals only.

Process



Kit Contents and Storage

cfKapture Kit - DX (1-2 mL) Catalog No.	CFK-DI5-2MLE	CFK-DI50-2MLE	Storage
Number of Preps	5	50	
CFL Buffer	12 mL	120 mL	15-25°C
CFW1 Buffer ¹	4 mL	40 mL	15-25°C
CFW2 Buffer ¹	2.5 mL	25 mL	15-25°C
Elution Buffer	550 µL	5.5 mL	15-25°C
Pro K Solution ²	220 µL	2.2 mL	2-8°C
MAG-CFB Particles ³	110 µL	1.1 mL	2-8°C

¹ Ethanol must be added prior to use. See Preparation of Reagents section.

² Once opened, reagents are usable until the expiration date on the product label. Be sure to close the lid firmly before storing reagents for later use.

Shipping and Storage

- ² Pro K Solution comes in a ready to use solution. Ships at room temperature. Store at 2-8°C.
- ³ MAG-CFB Particles ship at room temperature. Store at 2-8°C.

Safety Information

Any consumables, including plates, tubes, etc., used to process samples with infectious or microbial hazards should be disposed of in an appropriate biohazard waste bin. When working with chemicals, always wear a suitable lab coat, disposable gloves, and protective goggles. For more information, please consult the appropriate safety data sheets (SDSs). The SDS can be downloaded from the "Product Documents" tab when viewing this product at www.magbiogenomics.com.

Preparation of Reagents

Prepare the following components for each kit before use.

Equipment and Reagents to Be Supplied by the User

- Ethanol (100%)

Catalog No.	Component	Add 100% Ethanol	Storage
CFK-DI5-2MLE	CFW1 Buffer	5 mL*	15-25°C
	CFW2 Buffer	6.5 mL*	15-25°C

Catalog No.	Component	Add 100% Ethanol	Storage
CFK-DI50-2MLE	CFW1 Buffer	50 mL*	15-25°C
	CFW2 Buffer	65 mL*	15-25°C

*Ensure bottle/tube lid is closed tightly when preparing and storing reagents.

Protocol: 1 mL sample volume (15 mL tube format)


Equipment and Reagents to Be Supplied by the User

- RNase and DNase-free 1.5 mL or 2.0 mL microcentrifuge tubes
- 15 mL conical tubes
- Magnetic separation device for 1.5 mL/2.0 mL microcentrifuge tube (see page 8)
- Magnetic separation device for 15 mL conical tube (see page 8)
- 100% Ethanol
- Vortex
- Tube rotator
- Water bath, incubator, or heat block capable of 60°C

Things to do Before Starting

- Prepare all reagents according to the instructions on page 2
- Preheat water bath, incubator, or heat block to 60°C
- Warm Elution Buffer to 60°C


Protocol

 Bring the **MAG-CFB Particles** to room temperature for at least 30 min before use.

1. Transfer 1 mL of plasma to a 15 mL conical tube.
2. Add 40 µL of **Pro K Solution** and mix well by vortexing at maximum speed for 20 seconds.
3. Incubate the sample at 60°C for 10 minutes in a water bath. Mix by inverting the tube once during incubation.
4. Add 1 mL of **CFL Buffer** and mix well by vortexing at maximum speed for 60 seconds. Incubate the sample at room temperature for 5 minutes.
5. Add 750 µL of 100% Ethanol and 20 µL of **MAG-CFB Particles**. Mix immediately by vortexing at maximum speed for 20 seconds.

 *Complete resuspension of the **MAG-CFB Particles** is crucial for obtaining purity.*

6. Incubate the sample tube on a tube rotator for 20 minutes at room temperature at 10 rpm. Adjust the rotator angle to approximately 45 degrees for better mixing.
7. Remove the tube from rotator and place on a compatible 15 mL magnetic separation device to magnetize the **MAG-CFB Particles** for 20 minutes or until the magnetic particles are completely cleared from the solution.
8. With the tube on the magnetic separation device, remove and discard the cleared supernatant by pipetting.

 *Do not disturb the attracted beads while aspirating the supernatant.*
9. Remove the sample tube from the magnetic separation device.
10. Add 800 µL of **CFW1 Buffer** and resuspend the magnetic particles by vortexing at maximum speed for 1 minute or by pipetting up and down 10 times.
11. Transfer the solution from the 15 mL tube to a new 1.5 mL microcentrifuge tube.

12. Place the 1.5 mL sample tube on a compatible 1.5 mL magnetic separation device to magnetize the **MAG-CFB Particles** at room temperature for 5 minutes or until the magnetic particles are completely cleared from the solution.
13. Remove and discard the cleared supernatant.
⚠ Do not disturb the attracted beads while aspirating the supernatant.
14. Remove the sample from the magnetic separation device and repeat the wash by adding 800 μ L of **CFW1 Buffer** and resuspend the magnetic particles by vortexing at maximum speed for 1 minute or by pipetting up and down 10 times.
15. Place the 1.5 mL sample tube back on the magnetic separation device to magnetize the **MAG-CFB Particles** at room temperature for 5 minutes or until the magnetic particles are completely cleared from the solution.
16. With the sample still on the magnetic separation device, remove and discard the cleared supernatant.
⚠ Do not disturb the attracted beads while aspirating the supernatant.
17. Remove the sample tube from the magnetic separation device.
18. Add 800 μ L of **CFW2 Buffer** and resuspend the magnetic particles by vortexing at maximum speed for 1 minute or by pipetting up and down 10 times.
19. Place the 1.5 mL sample tube on the magnetic separation device to magnetize the **MAG-CFB Particles** at room temperature for 5 minutes or until the magnetic particles are completely cleared from the solution.
20. With the sample still on the magnetic separation device, remove and discard the cleared supernatant.
⚠ Do not disturb the attracted beads while aspirating the supernatant.
21. Repeat Steps 17-20 for a second **CFW2 Buffer** wash.
22. With the sample tube still on the magnetic separation device, air dry the **MAG-CFB Particles** for 10 minutes. Remove any residual liquid with a pipette.
⚠ It is critical to completely remove all liquid from the tube.
23. Remove the sample tube containing the **MAG-CFB Particles** from the magnetic separation device.
24. Add 50-100 μ L of **Elution Buffer** and completely resuspend the **MAG-CFB Particles** by vortexing at maximum speed for 10 seconds.
*⚠ Heat **Elution Buffer** at 60°C to improve the yield.*
25. Incubate at room temperature for 10 minutes.
26. Place the tubes back on the magnetic separation device and wait 5 minutes or until the magnetic particles are completely cleared from the **Elution Buffer**.
27. Transfer the clear supernatant containing the ccfDNA to a new 1.5 mL microcentrifuge tube and store at -20°C.

Protocol: 2 mL sample volume (15 mL tube format)

Equipment and Reagents to Be Supplied by the User

- RNase and DNase-free 1.5 mL or 2.0 mL microcentrifuge tubes
- 15 mL conical tubes
- Magnetic separation device for 1.5 mL/2.0 mL microcentrifuge tube (see page 8)
- Magnetic separation device for 15 mL conical tube (see page 8)
- 100% Ethanol
- Vortex
- Tube rotator
- Water bath, incubator, or heat block capable of 60°C

Things to do Before Starting


- Prepare all reagents according to the instructions on page 2
- Preheat water bath, incubator, or heat block to 60°C
- Warm Elution Buffer to 60°C

Protocol

 Bring the **MAG-CFB Particles** to room temperature for at least 30 min before use.

1. Transfer 2 mL of plasma to a 15 mL conical tube.
2. Add 40 µL of **Pro K Solution** and mix well by vortexing at maximum speed for 20 seconds.
3. Incubate the sample at 60°C for 10 minutes in a water bath. Mix by inverting the tube once during incubation.
4. Add 2 mL of **CFL Buffer** and mix well by vortexing at maximum speed for 60 seconds. Incubate the sample at room temperature for 10 minutes.
5. Add 1.5 mL of 100% Ethanol and 20 µL of **MAG-CFB Particles**. Mix immediately by vortexing at maximum speed for 20 seconds.

 *Complete resuspension of the **MAG-CFB Particles** is crucial for obtaining purity.*

6. Incubate sample tube on a tube rotator for 20 minutes at room temperature at 10 rpm. Adjust the rotator angle to approximately 45 degrees for better mixing.
 7. Remove the tube from rotator and place on a compatible 15 mL magnetic separation device to magnetize the **MAG-CFB Particles** for 20 minutes or until the magnetic particles are completely cleared from the solution.
 8. With the tube on the magnetic separation device, remove and discard the cleared supernatant by pipetting.
-  *Do not disturb the attracted beads while aspirating the supernatant.*
9. Remove the sample tube from the magnetic separation device.
 10. Add 800 µL of **CFW1 Buffer** and resuspend the magnetic particles by vortexing at maximum speed for 1 minute or by pipetting up and down 10 times.

11. Transfer all the solution from the 15 mL tube to a new 1.5 mL microcentrifuge tube.

12. Place the 1.5 mL sample tube on a compatible 2 mL magnetic separation device to magnetize the **MAG-CFB Particles** at room temperature for 5 minutes or until the magnetic particles are completely cleared from the solution.
13. Remove and discard the cleared supernatant.
⚠ Do not disturb the attracted beads while aspirating the supernatant.
14. Remove the sample from the magnetic separation device and repeat the wash by adding 800 µL of **CFW1 Buffer** and resuspend the magnetic particles by vortexing at maximum speed for 1 minute or by pipetting up and down 10 times.
15. Place the 1.5 mL sample tube back on the magnetic separation device to magnetize the **MAG-CFB Particles** at room temperature for 5 minutes or until the magnetic particles are completely cleared from the solution.
16. Remove and discard the cleared supernatant.
⚠ Do not disturb the attracted beads while aspirating the supernatant.
17. Remove the sample tube from the magnetic separation device.
18. Add 800 µL of **CFW2 Buffer** and resuspend the magnetic particles by vortexing at maximum speed for 1 minute or by pipetting up and down 10 times.
19. Place the 1.5 mL sample tube on the magnetic separation device to magnetize the **MAG-CFB Particles** at room temperature for 5 minutes or until the magnetic particles are completely cleared from the solution.
20. With the sample still on the magnetic separation device, remove and discard the cleared supernatant.
⚠ Do not disturb the attracted beads while aspirating the supernatant.
21. Repeat Steps 17-20 for a second **CFW2 Buffer** wash.
22. With the sample tube on the magnetic separation device, air dry the **MAG-CFB Particles** for 10 minutes. Remove any residual liquid with a pipette.
⚠ It is critical to completely remove all liquid from the tube.
23. Remove the sample tube containing the **MAG-CFB Particles** from the magnetic separation device.
24. Add 50-100 µL of **Elution Buffer** and completely resuspend the **MAG-CFB Particles** by vortexing at maximum speed for 10 seconds.
*⚠ Heat **Elution Buffer** at 60°C to improve the yield.*
25. Incubate at room temperature for 10 minutes.
26. Place the tubes back on the magnetic separation device and wait 5 minutes or until the magnetic particles are completely cleared from the **Elution Buffer**.
27. Transfer the clear supernatant containing the ccfDNA to a new 1.5 mL microcentrifuge tube and store at -20°C.

Troubleshooting Guide

Please use this guide to troubleshoot any problems that may arise. For further assistance, please contact technical support via:

Phone: US/Canada, +1 301-302-0144. Europe, +49 7250 33 13 403

Email: US/Canada, support@magbiogenomics.com. Europe, Info.europe@magbiogenomics.com

Symptoms	Possible Causes	Comments
Low yield of cfDNA	Incomplete resuspension of MAG-CFB Particles	Resuspend the MAG-CFB Particles by vortexing vigorously before use
	Loss of MAG-CFB Particles during operation	Avoid disturbing the MAG-CFB Particles during aspiration of supernatant
	Ethanol is not added into CFW1 Buffer or CFW2 Buffer	Add absolute 100% Ethanol to the CFW1 Buffer and CFW2 Buffer (see page 2 for instructions)
MAG-CFB Particles do not completely clear from the solution	Magnetizing time too short	Increase collection time on the magnet
Problems in downstream applications	Insufficient cfDNA in starting material	Use more starting material
	Ethanol carry-over	Dry the MAG-CFB Particles completely before elution
Carryover of MAG-CFB Particles	The eluate has particles and is not fully clear	Increase magnetization time. If small amount of carryover, place eluted sample on a magnetic separation device and perform an additional 5 min magnetization

Ordering Information

cfKapture Kit - DX

Catalog No.	Product	Description	Preps
CFK-DI50-400ULE	cfKapture Kit - DX (200-400 µL, 50 preps)	Purification of cell-free DNA (cfDNA) from 200-400 µL plasma	50
CFK-DI50-2MLE	cfKapture Kit - DX (1-2 mL, 50 preps)	Purification of cell-free DNA (cfDNA) from 1-2 mL plasma	50
CFK-DI50-5MLE	cfKapture Kit - DX (3-5 mL, 50 preps)	Purification of cell-free DNA (cfDNA) from 3-5 mL plasma	50

Related Products

HighPrep PCR - DX

Catalog No.	Product
AC-60005E	HighPrep PCR - DX (5 mL)
AC-60050E	HighPrep PCR - DX (50 mL)
AC-60250E	HighPrep PCR - DX (250 mL)
AC-60500E	HighPrep PCR - DX (500 mL)

Magnetic Separation Devices

Catalog No.	Product
MYMAG-96	Handheld Magnetic Separation Device (96 well microplate format)
MYMAG-96X	Magnetic Separation Device (96 well ring format)
MBMS-12	MagStrip Magnet Stand (1.5 mL x 12)
MBMS-31550	15 mL and 50 mL Magnetic Stand Combo (3 x 15 mL and 3 x 50 mL)



www.magbiogenomics.com